

TD2204B
Kit de Cariología (Preclínico)
Universidad Andrés Bello campus Viña del Mar

REF		ISO	Grano / Filos	 opt.	 max.
 801.314.010		806.314.001524.010	Medio (107 µm *)	160.000 min ⁻¹	450.000 min ⁻¹
 801.314.014		806.314.001524.014	Medio (107 µm *)	160.000 min ⁻¹	450.000 min ⁻¹
 801.314.018		806.314.001524.018	Medio (107 µm *)	160.000 min ⁻¹	450.000 min ⁻¹
 830L.314.012		806.314.234524.012	Medio (107 µm *)	160.000 min ⁻¹	450.000 min ⁻¹
 8368.314.023	 	806.314.257514.023	Fino (46 µm)	20.000 min ⁻¹	300.000 min ⁻¹
 H32.314.012	 	-	-	160.000 min ⁻¹	450.000 min ⁻¹
 4COMP.314.012	 	-	-	160.000 min ⁻¹	450.000 min ⁻¹
 830RM.314.009	 	-	Medio (107 µm *)	160.000 min ⁻¹	300.000 min ⁻¹
 889M.314.007	 	-	Medio (107 µm *)	160.000 min ⁻¹	300.000 min ⁻¹
 953AM.314.014	 	-	Medio (107 µm *)	160.000 min ⁻¹	160.000 min ⁻¹
 H1.204.018		500.204.001001.018	-	1.500 min ⁻¹	100.000 min ⁻¹
 H1.204.021		500.204.001001.021	-	1.500 min ⁻¹	100.000 min ⁻¹
 P1.204.018	  	-	-	2.000 min ⁻¹	8.000 min ⁻¹

* En algunos instrumentos el tamaño del grano puede variar en función de su forma y tamaño.