


TD2132F
Kit Prótesis Fija
Universidad Andrés Bello campus Viña del Mar

REF		ISO	Grano / Filos	 opt.	 max.
  6877K.314.014		806.314.297534.014	Grueso (151 µm *)	160.000 min ⁻¹	450.000 min ⁻¹
  6877K.314.016		806.314.297534.016	Grueso (151 µm *)	160.000 min ⁻¹	450.000 min ⁻¹
  6878K.314.016		806.314.298534.016	Grueso (151 µm *)	160.000 min ⁻¹	450.000 min ⁻¹
  6879K.314.018		806.314.299534.018	Grueso (151 µm *)	160.000 min ⁻¹	450.000 min ⁻¹
  6879K.314.021		806.314.299534.021	Grueso (151 µm *)	160.000 min ⁻¹	450.000 min ⁻¹
  8877K.314.014		806.314.297514.014	Fino (46 µm)	20.000 min ⁻¹	450.000 min ⁻¹
  8877K.314.016		806.314.297514.016	Fino (46 µm)	20.000 min ⁻¹	450.000 min ⁻¹
  8878K.314.016		806.314.298514.016	Fino (46 µm)	20.000 min ⁻¹	450.000 min ⁻¹
  8879K.314.018		806.314.299514.018	Fino (46 µm)	20.000 min ⁻¹	450.000 min ⁻¹
  8879K.314.021		806.314.299514.021	Fino (46 µm)	20.000 min ⁻¹	450.000 min ⁻¹
  6368.314.023		806.314.257534.023	Grueso (151 µm *)	160.000 min ⁻¹	300.000 min ⁻¹
  8368.314.023		806.314.257514.023	Fino (46 µm)	20.000 min ⁻¹	300.000 min ⁻¹
  8889.314.009		806.314.540514.009	Fino (46 µm)	20.000 min ⁻¹	300.000 min ⁻¹
  6850.314.012		806.314.199534.012	Grueso (151 µm *)	160.000 min ⁻¹	300.000 min ⁻¹
 811.314.033		806.314.038524.033	Medio (107 µm *)	100.000 min ⁻¹	100.000 min ⁻¹
  H284K.314.018		500.314.299072.018	12 Filos	20.000 min ⁻¹	300.000 min ⁻¹
 183L.204.070		330.204.682336.070	Peeso N°1	1.200 min ⁻¹	20.000 min ⁻¹
  183L.204.090		330.204.682336.090	Peeso N°2	1.200 min ⁻¹	20.000 min ⁻¹
  183L.204.110		330.204.682336.110	Peeso N°3	1.200 min ⁻¹	20.000 min ⁻¹
  17825.204.040		340.204.673458.040	Léntulo N°4	300 - 600 min ⁻¹	800 min ⁻¹

* En algunos instrumentos el tamaño del grano puede variar en función de su forma y tamaño.