

TD2134C
Kit Eliminación de Caries
Universidad Andrés Bello campus Viña del Mar

REF		ISO	Grano / Filos	 opt.	 max.
 ●● H34L.314.012		500.314.139008.012	-	160.000 min ⁻¹	450.000 min ⁻¹
 ● K1SM.204.018		-	-	1.500 min ⁻¹	40.000 min ⁻¹
 ● K1SM.204.021		-	-	1.500 min ⁻¹	40.000 min ⁻¹
 H1.204.010		500.204.001001.010	-	1.500 min ⁻¹	100.000 min ⁻¹
 H1.204.012		500.204.001001.012	-	1.500 min ⁻¹	100.000 min ⁻¹
 H1.204.016		500.204.001001.016	-	1.500 min ⁻¹	100.000 min ⁻¹
 H1.204.021		500.204.001001.021	-	1.500 min ⁻¹	100.000 min ⁻¹
 H1.204.027		500.204.001001.027	-	1.500 min ⁻¹	100.000 min ⁻¹