

TD3333A

Kit Preparaciones Estéticas Prof. Dr. A. Báez
Universidad Andrés Bello campus Viña del Mar

REF	ISO	Grano / Filos	 opt.	 max.
 6877K.314.014		806.314.297534.014 Grueso (151 μ m *)	160.000 min ⁻¹	450.000 min ⁻¹
 6877K.314.016		806.314.297534.016 Grueso (151 μ m *)	160.000 min ⁻¹	450.000 min ⁻¹
 6878K.314.016		806.314.298534.016 Grueso (151 μ m *)	160.000 min ⁻¹	450.000 min ⁻¹
 6879K.314.018		806.314.299534.018 Grueso (151 μ m *)	160.000 min ⁻¹	450.000 min ⁻¹
 6879K.314.021		806.314.299534.021 Grueso (151 μ m *)	160.000 min ⁻¹	450.000 min ⁻¹
 8877K.314.014		806.314.297514.014 Fino (46 μ m)	20.000 min ⁻¹	450.000 min ⁻¹
 8877K.314.016		806.314.297514.016 Fino (46 μ m)	20.000 min ⁻¹	450.000 min ⁻¹
 8878K.314.016		806.314.298514.016 Fino (46 μ m)	20.000 min ⁻¹	450.000 min ⁻¹
 8879K.314.018		806.314.299514.018 Fino (46 μ m)	20.000 min ⁻¹	450.000 min ⁻¹
 8879K.314.021		806.314.299514.021 Fino (46 μ m)	20.000 min ⁻¹	450.000 min ⁻¹
 6368.314.023		806.314.257534.023 Grueso (151 μ m *)	160.000 min ⁻¹	300.000 min ⁻¹
 8368.314.023		806.314.257514.023 Fino (46 μ m)	20.000 min ⁻¹	300.000 min ⁻¹
 6837KR.314.014		806.314.158534.014 Grueso (151 μ m *)	160.000 min ⁻¹	450.000 min ⁻¹
 8837KR.314.014		806.314.158514.014 Fino (46 μ m)	20.000 min ⁻¹	450.000 min ⁻¹
 8889.314.009		806.314.540514.009 Fino (46 μ m)	20.000 min ⁻¹	300.000 min ⁻¹
 6801.314.016		806.314.001534.016 Grueso (151 μ m *)	160.000 min ⁻¹	450.000 min ⁻¹
 6850.314.012		806.314.199534.012 Grueso (151 μ m *)	160.000 min ⁻¹	300.000 min ⁻¹
 811.314.033		806.314.038524.033 Medio (107 μ m *)	100.000 min ⁻¹	100.000 min ⁻¹
 811.314.037		806.314.039524.037 Medio (107 μ m *)	100.000 min ⁻¹	100.000 min ⁻¹
 H283K.314.016		500.314.298072.016 10 Filos	20.000 min ⁻¹	450.000 min ⁻¹
 H283K.314.021		500.314.298072.021 12 Filos	20.000 min ⁻¹	300.000 min ⁻¹
 H284K.314.018		500.314.299072.018 12 Filos	20.000 min ⁻¹	300.000 min ⁻¹

* En algunos instrumentos el tamaño del grano puede variar en función de su forma y tamaño.