

TD2198D

Kit Preparaciones Estéticas Prof. Dr. A. Báez
Universidad Andrés Bello campus Viña del Mar

REF		ISO	Grano / Filos	 opt.	 max.
  ● 6877K.314.014		806.314.297534.014	Grueso (151 μ m *)	160.000 min ⁻¹	450.000 min ⁻¹
  ● 6877K.314.016		806.314.297534.016	Grueso (151 μ m *)	160.000 min ⁻¹	450.000 min ⁻¹
  ● 6878K.314.016		806.314.298534.016	Grueso (151 μ m *)	160.000 min ⁻¹	450.000 min ⁻¹
  ● 6879K.314.018		806.314.299534.018	Grueso (151 μ m *)	160.000 min ⁻¹	450.000 min ⁻¹
  ● 6879K.314.021		806.314.299534.021	Grueso (151 μ m *)	160.000 min ⁻¹	450.000 min ⁻¹
  ● 8877K.314.014		806.314.297514.014	Fino (46 μ m)	20.000 min ⁻¹	450.000 min ⁻¹
  ● 8877K.314.016		806.314.297514.016	Fino (46 μ m)	20.000 min ⁻¹	450.000 min ⁻¹
  ● 8878K.314.016		806.314.298514.016	Fino (46 μ m)	20.000 min ⁻¹	450.000 min ⁻¹
  ● 8879K.314.018		806.314.299514.018	Fino (46 μ m)	20.000 min ⁻¹	450.000 min ⁻¹
  ● 8879K.314.021		806.314.299514.021	Fino (46 μ m)	20.000 min ⁻¹	450.000 min ⁻¹
  ● 6368.314.023		806.314.257534.023	Grueso (151 μ m *)	160.000 min ⁻¹	300.000 min ⁻¹
  ● 8368.314.023		806.314.257514.023	Fino (46 μ m)	20.000 min ⁻¹	300.000 min ⁻¹
  ● 6837KR.314.014		806.314.158534.014	Grueso (151 μ m *)	160.000 min ⁻¹	450.000 min ⁻¹
  ● 8837KR.314.014		806.314.158514.014	Fino (46 μ m)	20.000 min ⁻¹	450.000 min ⁻¹
  ● 8889.314.009		806.314.540514.009	Fino (46 μ m)	20.000 min ⁻¹	300.000 min ⁻¹
  ● 6801.314.016		806.314.001534.016	Grueso (151 μ m *)	160.000 min ⁻¹	450.000 min ⁻¹
  ● 6850.314.012		806.314.199534.012	Grueso (151 μ m *)	160.000 min ⁻¹	300.000 min ⁻¹
  811.314.033		806.314.038524.033	Medio (107 μ m *)	100.000 min ⁻¹	100.000 min ⁻¹
  ● H283K.314.016		500.314.298072.016	10 Filos	20.000 min ⁻¹	450.000 min ⁻¹
  ● H283K.314.021		500.314.298072.021	12 Filos	20.000 min ⁻¹	300.000 min ⁻¹
  ● H284K.314.018		500.314.299072.018	12 Filos	20.000 min ⁻¹	300.000 min ⁻¹

* En algunos instrumentos el tamaño del grano puede variar en función de su forma y tamaño.